

# Technical Bulletin 173

Version 1.0      22<sup>nd</sup> November 2016

---

## Updated Tiger and Tiger Select firmware and PC software

---

The Tiger and Tiger Select firmware and PC software has been updated with a number of enhancements. Please install the new versions below to ensure you are up to date.

- Tiger & Tiger Select firmware version: V0.7.05
- Tiger & Tiger Select software version: V1.0.0.89

The improvements that have been made within this release are detailed below.

### Calibration routine

The zero, span 1 and span 2 countdown times within the calibration routine have now been extended to 30 seconds.

### Electrode stack options

The electrode stack icons within the configuration screen have now been removed. The lamp settings should be configured via the "lamp" section of Tiger PC.

### Run time

The total instrument run time is displayed within the information icon on screen 2 (Tiger), and screen 5 (Select).

### System Requirements

Tiger PC software must be used in conjunction with a PC or laptop using Windows Vista, Windows 7 or Windows 10.

### Gas table prompt

When a new version of gas table becomes available an alert will appear in Tiger PC. The user can accept or reject the gas table update. Ion Science however, advises that users ensure the gas tables used within the Tiger are current so they reflect the most up to date information.

The firmware and PC software can be upgraded directly via your existing version of Tiger PC. Please refer to the relevant sections of the user manual for guidance. Alternatively, Tiger PC can be downloaded from the Ion Science website:

<http://www.ionscience.com/customer-support/media-centre/instrument-software/>

### Gas table update

The gas table information has been updated in line with this change and is reflected within the:

- Tiger and Tiger Select instrument gas table
- Online Gas Search
- Technical Application Article TA-02 Ion Science PID response factors

### For more information please contact us

Email: [marketing@ionscience.com](mailto:marketing@ionscience.com)

Tel: +44 (0)1763 208503